

CYSTATIN C CONTROL - LEVEL 3 (CYSC CONTROL 3)

CAT. NO. CYS 5020

LOT NO. I305CY

EXPIRY: 2021-12-28

SIZE: 3 x 2 ml

GTIN: 05055273207422

SAMPLE PREPARATION

The control is ready for use.

STORAGE AND STABILITY

UNOPENED: Material is stable to expiry date when stored at +2°C to +8°C, if kept capped in original container and free from contamination.

OPENED: Material is stable for 30 days at +2°C to +8°C in opened vials. Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

VALUE

3.39 mg/l 3.05 - 3.73 mg/l

SAFETY PRECAUTIONS

This Control material is derived from human serum obtained from volunteer donors. All donors have been found negative for Hepatitis B surface antigen and Anti-HIV antibody. However, since no test method can offer complete assurance that products will not transmit infectious agents, it is recommended that this produce is handled with the same precautions used for patient samples.

For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

This material contains Sodium Azide. Avoid ingestion or contact with skin or mucous membranes. In case of skin contact, flush affected area with copious amounts of water. In case of contact with eyes or if ingested, seek immediate medical attention.

Sodium Azide reacts with lead and copper plumbing, to form potentially explosive azides. When disposing of such reagents, flush with large volumes of water to prevent azide build up. Exposed metal surfaces should be cleaned with 10% sodium hydroxide.

28 Feb '20 ne

THIS PAGE IS INTENTIONALLY BLANK