

PRODUCT INFORMATION

IAS3113

1943EC

Please note targets and ranges are currently not available for 1,25 OH Vitamin D. These will be updated in due course.

CCS6438

IMMUNOASSAY SPECIALITY I - LEVEL I (IA SPECIALITY I LEV I)

CAT NO.	IAS3113	LOT NO.	1943EC
SIZE:	5 x 2 ml	EXPIRY:	2021-11-28
GTIN:	05055273207293		

INTENDED USE

This product is intended for *in vitro* diagnostic use, in the quality control of the accuracy of Immunoassays on clinical chemistry systems. This material can be used to monitor the control of accuracy or the control of reproducibility of immunoassays.

SAFETY PRECAUTIONS AND WARNINGS

For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests.

However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

Health and Safety Data Sheets are available on request.

STORAGE AND STABILITY

OPENED: Store refrigerated (+2°C to +8°C). Reconstituted serum is stable for 5 days at +2°C to +8°C if kept capped in original container and free from contamination, or 4 weeks frozen once at -20°C. Anti-TG is stable for 3 days at +2°C to +8°C. C-Peptide, IGF-I and Procalcitonin are stable for 1 day at +2°C to +8°C. Osteocalcin is stable for 4 hours at +2°C to +8°C. Parathyroid Hormone (PTH) should be tested within 4 hours of reconstitution when stored at +2°C to +8°C, or within 2 weeks when stored below -20°C.

Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

UNOPENED: Store refrigerated (+2°C to +8°C). Stable to expiration date printed on individual vials.

Bacterial contamination of the reconstituted serum will cause reductions in the stability of many components. If bacterial contamination is suspected, the vial should be discarded and a fresh vial reconstituted.

PREPARATION

The Immunoassay Speciality I Control is supplied lyophilised.

- Carefully reconstitute each vial of lyophilised serum with exactly 2 ml of distilled water at +15°C to +25°C. Close the bottle and allow to stand for 30 minutes before use. Ensure contents are completely dissolved by swirling gently. Avoid formation of foam. Do not shake.
- Refer to the Control section of the individual analyser application.
- Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

MATERIALS PROVIDED

Immunoassay Speciality I - Level I 5 x 2 ml

MATERIAL REQUIRED BUT NOT PROVIDED

Volumetric pipette

VALUE ASSIGNMENT

Each batch of Immunoassay Speciality I Control is submitted to a number of reference laboratories and values are assigned from a consensus of results obtained by these laboratories, using a unique statistical analysis. With each batch, a control range is provided for individual parameters and each parameter method. The control range is equivalent to the assigned mean ± 2 S.D. This results in extremely accurate values, which may be confidently used by laboratories to ensure the accuracy of their methods.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 9445 1070 or email Technical.Services@randox.com.

20 May 20 pl

IMMUNOASSAY SPECIALITY I - LEVEL 1 (IA SPECIALITY I LEV 1)

Cat. No. IAS3113 Lot. No. 1943EC Size 5 x 2 ml Expiry 2021-11-28

Analyte	unit	Target	Range		methods
			low	high	
25-OH Vitamin D	nmol/l	51.4	38.6	64.3	Diasorin Liaison
	nmol/l	33.2	24.9	41.5	Abbott Architect
	nmol/l	76.3	57.2	95.4	Siemens Centaur XP/XPT/Classic
	nmol/l	39.7	29.8	49.6	BioMerieux Vidas
	nmol/l	43.2	32.4	54.0	Beckman Dxl 600/800
	nmol/l	33.5	25.1	41.9	Abbott Architect (5P02)/ Alinity (8P45)
	nmol/l	61.2	45.9	76.5	Diasorin Liaison XL
	nmol/l	39.4	29.6	49.3	Roche Vitamin D Total II
	nmol/l	60.6	45.5	75.8	Siemens Atellica IM
Anti TG	kU/l	94.8	71.1	119	Siemens Immulite 2000/2500
	kU/l	64.2	48.2	80.3	Abbott Architect
	kU/l	447	335	559	Roche Cobas E411
	kU/l	401	301	501	Roche Cobas 6000/8000
	kU/l	202	152	253	Beckman Dxl 600/800
	kU/l	355	266	444	Siemens Centaur XP/XPT/Classic
	kU/l	238	179	298	BioMerieux Vidas
	kU/l	370	278	463	Siemens Atellica IM
Anti TPO	kU/l	40.2	30.2	50.3	Abbott Architect
	kU/l	42.8	32.1	53.5	Siemens Immulite 2000/2500
	kU/l	11.6	8.70	14.5	Roche Cobas E411
	kU/l	30.8	23.1	38.5	Beckman Dxl 600/800
	kU/l	144	108	180	Siemens Centaur XP/XPT/Classic
	kU/l	43.4	32.6	54.3	BioMerieux Vidas
	kU/l	135	101	169	Siemens Atellica IM
C-Peptide	nmol/l	0.322	0.242	0.403	Siemens Centaur XP/XPT/Classic
	ng/ml	0.972	0.731	1.21	
	nmol/l	0.487	0.365	0.609	Roche Cobas 6000/8000
	ng/ml	1.47	1.10	1.84	
	nmol/l	0.506	0.380	0.633	Diasorin Liaison
	ng/ml	1.53	1.15	1.91	
	nmol/l	0.377	0.283	0.471	Abbott Architect
	ng/ml	1.14	0.854	1.43	
	nmol/l	0.326	0.245	0.408	Siemens Immulite 2000/2500
	ng/ml	0.984	0.740	1.23	
	nmol/l	0.494	0.371	0.618	Roche Cobas E411
	ng/ml	1.49	1.12	1.86	
	nmol/l	1.150	0.863	1.438	Ortho Vitros ECi
	ng/ml	3.47	2.61	4.34	
IGF 1	µg/l	78.8	59.1	98.5	Siemens Immulite 2000/2500
	µg/l	68.9	51.7	86.1	Diasorin Liaison
	µg/l	48.5	36.4	60.6	Roche Cobas 6000/8000
Insulin	mU/l	4.70	3.53	5.88	Abbott Architect
	µU/ml	4.70	3.53	5.87	

IMMUNOASSAY SPECIALITY I - LEVEL 1 (IA SPECIALITY I LEV 1)

Cat. No. IAS3113 Lot. No. 1943EC

Size 5 x 2 ml Expiry 2021-11-28

Analyte	unit	Target	Range		methods
			low	high	
Insulin	mU/l	5.31	3.98	6.64	Roche Cobas 6000/8000
	µU/ml	5.31	3.98	6.64	
	mU/l	4.96	3.72	6.20	Siemens Centaur XP/XPT/Classic
	µU/ml	4.96	3.72	6.20	
	mU/l	3.84	2.88	4.80	Beckman Dxl 600/800
	µU/ml	3.84	2.88	4.80	
	mU/l	5.18	3.89	6.48	Roche Cobas E411
	µU/ml	5.18	3.89	6.47	
Osteocalcin	µg/l	16.5	12.4	20.6	Siemens Immulite 2000/2500
	µg/l	25.5	19.1	31.9	Roche Cobas 6000/8000
	µg/l	24.7	18.5	30.9	Roche Cobas E411
	µg/l	25.2	18.9	31.5	Diasorin Liaison
Parathyroid Hormone (PTH)	pmol/l	5.00	3.75	6.25	Abbott Architect
	pg/ml	47.5	35.6	59.4	
	pmol/l	3.30	2.48	4.13	Beckman Dxl 600/800
	pg/ml	31.3	23.6	39.0	
	pmol/l	5.38	4.04	6.73	Siemens Immulite 2000/2500
	pg/ml	51.1	38.4	63.8	
	pmol/l	3.59	2.69	4.49	Roche Cobas 6000/8000
	pg/ml	34.1	25.5	42.7	
	pmol/l	3.98	2.99	4.98	Roche Cobas 4000/e411 PTH/PTH STAT
	pg/ml	37.8	28.4	47.2	
Procalcitonin	µg/l	0.780	0.585	0.975	Brahms Kryptor
	µg/l	1.32	0.990	1.65	BioMerieux Vidas
	µg/l	0.850	0.638	1.06	Roche Elecsys/Cobas/Modular
	µg/l	1.69	1.27	2.11	Diasorin IRMA
	µg/l	1.20	0.900	1.50	Siemens Centaur XP/XPT/Classic
	µg/l	1.23	0.923	1.54	Siemens Centaur CP
	µg/l	2.11	1.48	2.74	Diazyme PCT
	µg/l	0.902	0.677	1.13	Abbott Architect Brahms PCT
	µg/l	2.22	1.22	3.22	Nexus IB10
	µg/l	0.908	0.625	1.19	Ortho Vitros Brahms PCT