

GLUTAMINE CONTROL - LEVEL I

CAT. NO. GM1376

LOT NUMBER: 257GM

EXPIRY: 2017-04

INTENDED USE

This product is intended for use in the control of Glutamine on RX **daytona** clinical chemistry systems.

FOR BIOTECHNOLOGY INDUSTRIAL USE. Not for use in diagnostic procedures.

SAFETY PRECAUTIONS AND WARNINGS

For biotechnology industrial use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

Health and Safety Data Sheets are available on request.

STORAGE AND STABILITY

Unopened: Store refrigerated (+2°C to +8°C). Stable to expiration date printed on the individual vials.

Carefully reconstitute the contents of one vial with **5 ml** of distilled water and allow to stand for 30 minutes. Mix carefully to ensure homogeneity of the solution. After reconstitution, the material is stable for 7 days at +2 to +8°C.

Please note: these controls should not be used if exposed to temperatures above +25°C for greater than 24 hours, as their accuracy may be affected.

PREPARATION FOR USE

1. Mix by inversion at least 5 times to ensure homogeneity prior to use.
2. Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

MATERIALS PROVIDED

Glutamine Control Level I (5 x 5 ml)

MATERIALS REQUIRED BUT NOT PROVIDED

None

ASSIGNED VALUES

A value (mean) has been assigned to this lot of Control at Randox Laboratories Ltd.

ANALYTE	TARGET VALUE	RANGE
GLUTAMINE	1.07 mmol/l	0.91 - 1.23 mmol/l

27 Apr '16 ne

THIS PAGE IS INTENTIONALLY BLANK