

## MICROALBUMIN LIQUID CONTROLS (mALB CONTROL)

**CAT. NO.** MA 1361

**SIZE:** 6 x 1 ml

**LOT NOS.** 1057MA - 1058MA

**EXPIRY:** 2023-02-28

**GTIN:** 05055273204513

### INTENDED USE

The Microalbumin controls are intended for *in vitro* diagnostic use, in the quality control of the Microalbumin assay on clinical chemistry systems.

### SAMPLE PREPARATION

The Controls are ready for use.

### STABILITY

The Microalbumin controls are supplied ready for use, and are stable up to the expiry date when capped and stored at +2°C to +8°C in the absence of contamination. Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

### LEVELS

I.T. METHOD	MEAN VALUE mg/l	RANGE mg/l
1057MA	33.7	27.0 - 40.5
1058MA	174.0	139.2 - 208.8

### SAFETY

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

N.B. Please note that these controls have been developed for use with Randox Reagents and may not be suitable for use with material from other commercial suppliers.

**EC REP** Randox Teoranta, Meenmore,  
Dungloe, Donegal,  
F94 TV06, Ireland

(W/H: 52) 16 Apr '21 ne

**THIS PAGE IS INTENTIONALLY BLANK**