

LIPID CONTROL - LEVEL 2 (LPD CONTROL 2)

Cat. No. LE2669 **Lot No.** 2969CH
Size: 5 x 1 ml **Expiry:** 2025-03-28
GTIN: 05055273204193

INTENDED USE

This product is intended for *in vitro* use in the quality control of Direct HDL, Direct LDL, Lipoprotein (a), Apolipoprotein A-I, Apolipoprotein B, Cholesterol and Triglyceride methods on clinical chemistry systems.

SAFETY PRECAUTIONS AND WARNINGS

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV1 & HIV2) antibody, Hepatitis B surface antigen (HbsAg) and the Hepatitis C virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests. However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting disease. For *in vitro* diagnostic use only.

STORAGE AND STABILITY

Unopened Lipid Control is stable until the expiry date printed on the product label when stored between +2°C and +8°C. Once reconstituted, the components of the serum are stable for 7 days at +2°C to +8°C, and 4 weeks at -20°C when frozen once. The following exceptions apply: LP(a) is stable for 16 weeks at -20°C when frozen once. Values may drop by up to 10% for Direct LDL Cholesterol when stored for 4 weeks at -20°C.

PREPARATION FOR USE

Open the vial carefully, avoiding any loss of the material and reconstitute with 1 ml of distilled water. Replace the rubber stopper, close the vial and leave to stand for 30 minutes before use. Ensure that all traces of dry material are dissolved by swirling gently.

MATERIALS PROVIDED

Lipid Control - Level 2 5 x 1 ml

MATERIALS REQUIRED BUT NOT PROVIDED

Distilled water
Volumetric pipette

VALUE ASSIGNMENT

Due to the variation caused by test equipment, test reagents and laboratory technique, the quoted ranges are provided for guidance. It is recommended that these ranges are used until each laboratory has established its own ranges, based on individual laboratory requirements.

Each batch of Lipid Control is submitted to a number of external laboratories. Values are assigned from a consensus of results obtained by these laboratories and internal testing conducted at Randox Laboratories Ltd.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 9445 1070 or email Technical.Services@randox.com.



Randox Teoranta, Meenmore,
Dungloe, Donegal,
F94 TV06, Ireland

31 Aug '22 me

LIPID CONTROL LEVEL 2 (LPD CONTROL 2)

Lot. No. 2969CH Cat. No. LE2669

Size 5 x 1ml Expiry 2025-03-28

Range

Analyte	unit	Target	low	high	methods
Apolipoprotein A-1	g/l	1.56	1.28	1.84	Immunoturbidimetric
	mg/dl	156	128	184	
	g/l	1.61	1.32	1.90	Nephelometric
	mg/dl	161	132	190	
Apolipoprotein B	g/l	1.16	0.95	1.37	Immunoturbidimetric
	mg/dl	116	95.1	137	
	g/l	1.17	0.96	1.38	Nephelometric
	mg/dl	117	95.9	138	
Cholesterol	mmol/l	5.82	5.06	6.58	Cholesterol Oxidase - Abell Kendall
	mg/dl	225	195	255	
	mmol/l	5.69	4.95	6.43	Siemens Dimension
	mg/dl	220	191	249	
	mmol/l	5.86	5.10	6.62	Cholesterol Oxidase - IDMS
	mg/dl	226	197	255	
HDL - Cholesterol	mmol/l	1.44	1.22	1.66	Direct Clearance Method
	mg/dl	55.6	47.1	64.1	
	mmol/l	1.23	1.05	1.41	Phosphotungstic acid pptn.
	mg/dl	47.5	40.5	54.5	
	mmol/l	1.36	1.16	1.56	Direct HDL Immunoseparation
	mg/dl	52.5	44.8	60.2	
	mmol/l	1.41	1.20	1.62	Direct HDL PEGME
	mg/dl	54.4	46.3	62.5	
	mmol/l	1.58	1.34	1.82	Direct HDL PPD
	mg/dl	61.0	51.7	70.3	
LDL - Cholesterol	mmol/l	1.32	1.12	1.52	Direct HDL Roche 4th Generation
	mg/dl	51.0	43.2	58.8	
	mmol/l	1.57	1.33	1.81	HDL - Ultra
	mg/dl	60.6	51.3	69.9	
LDL - Cholesterol	mmol/l	3.72	3.16	4.28	Direct Clearance Method
	mg/dl	144	122	166	
	mmol/l	3.30	2.81	3.80	Calculated
	mg/dl	127	108	146	
Lipoprotein (a)	mmol/l	3.19	2.71	3.67	Selective detergent methods
	mg/dl	123	105	141	
Lipoprotein (a)	mg/dl	24.4	19.5	29.3	Immunoturbidimetric
	nmol/l	45.2	36.2	54.2	
Triglycerides	mmol/l	2.28	1.92	2.64	Lipase/GPO-PAP no correction
	mg/dl	202	170	234	
	mmol/l	2.38	2.00	2.76	Lipase/GK UV no correction
	mg/dl	211	177	245	

LIPID CONTROL LEVEL 2 (LPD CONTROL 2)

Lot. No. 2969CH Cat. No. LE2669

Size 5 x 1ml Expiry 2025-03-28

Range

Analyte	unit	Target	low	high	methods
Triglycerides	mmol/l	2.32	1.95	2.69	Lipase/Glycerol Dehydrogenase
	mg/dl	205	173	237	