

IMMUNOASSAY SPECIALITY II - LEVEL 2 (IA SPECIALITY II LEV 2)

CAT. NO.	IAS 3118	LOT NO.	I569EC
SIZE:	5 x 1ml	EXPIRY:	2017-10-28
GTIN:	05055273207330		

INTENDED USE

This product is intended for *in vitro* diagnostic use in the quality control of the accuracy of Immunoassays on clinical chemistry systems. This material can be used to monitor the control of accuracy or the control of reproducibility of immunoassays.

SAFETY PRECAUTIONS AND WARNINGS

For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

Human source material from which this product has been derived has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests.

However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

Health and Safety Data Sheets are available on request.

STORAGE AND STABILITY

OPENED: Store refrigerated (+2°C to +8°C). In reconstituted serum Renin is stable for 5 days, Procalcitonin is stable for 1 day, Gastrin and Calcitonin are stable for 8 hours at +2°C to +8°C if kept capped in original container and free from contamination. The control is stable if frozen once for 28 days at -20°C. Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

UNOPENED: Store refrigerated (+2°C to +8°C). Stable to expiration date printed on individual vials.

Bacterial contamination of the reconstituted serum will cause reductions in the stability of many components. If bacterial contamination is suspected, the vial should be discarded and a fresh vial reconstituted.

PREPARATION

Immunoassay Speciality II is supplied lyophilised.

1. Carefully reconstitute each vial of lyophilised serum with exactly 1ml of distilled water at +20°C to +25°C. Close bottle and allow to stand for 30 minutes before use. Ensure contents are completely dissolved by swirling gently. Avoid formation of foam. Do not shake.
2. Refer to the Control section of the individual analyser application.
3. Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

MATERIALS PROVIDED

Immunoassay Speciality II - Level 2 5 x 1ml

MATERIAL REQUIRED BUT NOT PROVIDED

Volumetric Pipette

VALUE ASSIGNMENT

Each batch of Immunoassay Speciality II is submitted to a number of reference laboratories and values are assigned from a consensus of results obtained by these laboratories using a unique statistical analysis. With each batch, a control range is provided for individual parameters and each parameter method. The control range is equivalent to the assigned mean ± 2 S.D. This results in extremely accurate values, which may be confidently used by laboratories to ensure the accuracy of their methods.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 94451070 or email Technical.Services@randox.com

12 Aug 16 pl

IMMUNOASSAY SPECIALITY II - LEVEL 2 (IA SPECIALITY LEV 2)

Cat. No. IAS3118 Lot No. 1569EC Size: 5 x 1ml Expiry: 2017-10-28

Range					
Analyte	unit	Target	low	high	methods
Calcitonin	pmol/l	29.2	21.9	36.5	Siemens Immulite 2000/2500
	pg/ml	100	74.7	125	
	pmol/l	22.2	16.7	27.8	Diasorin Liaison
	pg/ml	75.8	56.9	94.7	
	pmol/l	35.2	26.4	44.0	Roche Elecsys/Cobas/Modular
	pg/ml	120	90.0	150	
Gastrin	pmol/l	87.2	65.4	109	Siemens Immulite 2000/2500
	pg/ml	182	136	228	
	pmol/l	40.8	30.6	51.0	CISBIO RIA
	pg/ml	85.0	63.7	106	
Plasma Renin Activity	µg/l/h	2.63	1.84	3.42	Diasorin/RRENCTK R Angiotensin RIA
Procalcitonin	µg/l	0.593	0.445	0.741	Brahms Kryptor
	µg/l	2.34	1.76	2.93	BioMerieux Vidas
	µg/l	1.16	0.870	1.45	Roche Elecsys/Cobas/Modular
	µg/l	2.65	1.99	3.31	Siemens Centaur XP/XPT/Classic
Renin	mlU/l	61.5	46.1	76.9	Diasorin Liason direct Renin
	pg/ml	36.2	27.1	45.3	
	mlU/l	41.1	30.8	51.3	CISBIO RIA
	pg/ml	24.1	18.1	30.1	
	mlU/l	45.0	33.8	56.3	IDS iSYS
	pg/ml	26.5	19.9	33.1	