

## IMMUNOASSAY SPECIALITY II - LEVEL 2 (IA SPECIALITY II LEV 2)

**CAT. NO.** IAS 3118 **LOT NO.** 2288EC  
**SIZE** 5 x 1 ml **EXPIRY:** 2025-04-28  
**GTIN:** 05055273207330

### INTENDED USE

This product is intended for *in vitro* diagnostic use, in the quality control of the accuracy of Immunoassays on clinical chemistry systems. This material can be used to monitor the control of accuracy or the control of reproducibility of immunoassays.

### SAFETY PRECAUTIONS AND WARNINGS

For *in vitro* diagnostic use only. Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV 1, HIV 2) antibody, Hepatitis B Surface Antigen (HbsAg), and Hepatitis C Virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests.

However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting infectious diseases and disposed of accordingly.

Health and Safety Data Sheets are available on request.

### STORAGE AND STABILITY

Unopened: Unreconstituted control is stable up to the expiry date shown on the side of each individual bottle.

Opened: Stability claims once reconstituted are listed below.

Analyte	+2°C to +8°C	Aliquoted and frozen once at -20°C
Gastrin	8 hours	28 days
Calcitonin	8 hours	28 days
Siemens ADVIA Centaur CALCT Assay (Calcitonin)	8 hours	14 days
Procalcitonin	1 day	28 days
Renin	5 days	28 days

Only the required amount of product should be removed. After use, any residual product should NOT BE RETURNED to the original vial.

Bacterial contamination of the reconstituted serum will cause reductions in the stability of many components. If bacterial contamination is suspected, the vial should be discarded and a fresh vial reconstituted.

### PREPARATION

Immunoassay Speciality II is supplied lyophilised.

- Carefully reconstitute each vial of lyophilised serum with exactly 1 ml of distilled water at +15°C to +25°C. Close the bottle and allow to stand for 30 minutes before use. Ensure contents are completely dissolved by swirling gently. Avoid formation of foam. Do not shake.
- Refer to the Control section of the individual analyser application.
- Refrigerate any unused material. Prior to reuse, mix contents thoroughly.

### MATERIALS PROVIDED

Immunoassay Speciality II - Level 2 5 x 1 ml

### MATERIAL REQUIRED BUT NOT PROVIDED

Volumetric pipette

### VALUE ASSIGNMENT

Due to the variation caused by test equipment, test reagents and laboratory technique, the quoted ranges are provided for guidance. It is recommended that these ranges are used until each laboratory has established its own ranges, based on individual laboratory requirements.

Each batch of Immunoassay Speciality II is submitted to a number of reference laboratories and values are assigned from a consensus of results obtained by these laboratories, using a unique statistical analysis. With each batch, a control range is provided for individual parameters and each parameter method. The control range is equivalent to the assigned mean  $\pm 2$  S.D.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 9445 1070 or email [Technical.Services@randox.com](mailto:Technical.Services@randox.com).

EC	REP
----	-----

Randox Teoranta, Meenmore,  
Dungloe, Donegal,  
F94 TV06, Ireland

28 Sep '23 me

## IMMUNOASSAY SPECIALITY II - LEVEL 2 (IA SPECIALITY II LEV 2)

Cat. No. IAS3118 Lot. No. 2288EC

Size 5 x 1 ml Expiry 2025-04-28

Range					
Analyte	unit	Target	low	high	methods
Calcitonin	pmol/l	29.3	22.0	36.6	Siemens Immulite 2000/2500
	pg/ml	100	75.0	125	
	pmol/l	31.0	23.3	38.8	Roche Elecsys/Cobas/Modular
	pg/ml	106	79.5	133	
	pmol/l	27.9	20.9	34.9	Diasorin Liaison XL
	pg/ml	95.1	71.3	119	
	pmol/l	32.2	24.2	40.3	Roche Cobas e402/e801
	pg/ml	110	82.5	138	
Gastrin	pmol/l	73.3	55.0	91.6	Siemens Immulite 2000/2500
	pg/ml	153	115	191	
Procalcitonin	µg/l	4.41	3.31	5.51	BioMerieux Vidas
	µg/l	2.55	1.91	3.19	Roche Elecsys/Cobas/Modular
	µg/l	2.47	1.85	3.09	Abbott Architect / Alinity Brahms PCT
	µg/l	2.49	1.87	3.11	Roche Cobas e402/e801
	µg/l	3.57	2.68	4.46	Beckman Access PCT
	µg/l	3.90	2.93	4.88	Siemens Atellica IM 11202699
Renin	mIU/l	37.4	28.1	46.8	Diasorin Liaison direct Renin
	pg/ml	22.0	16.5	27.5	
	mIU/l	38.4	28.8	48.0	Diasorin Liaison XL direct Renin
	pg/ml	22.6	16.9	28.3	