

LIPID CONTROL - LEVEL I (LPD CONTROL I)

Cat. No. LE2661 **Lot No.** 2392CH **Size:** 5 x 3 ml **Expiry:** 2018-10

INTENDED USE

This product is intended for *in vitro* use, in the quality control of Direct HDL, Direct LDL, Lipoprotein (a), Apolipoprotein A-I, Apolipoprotein B, Cholesterol and Triglyceride methods on clinical chemistry systems.

SAFETY PRECAUTIONS AND WARNINGS

Human source material, from which this product has been derived, has been tested at donor level for the Human Immunodeficiency Virus (HIVI & HIV2) antibody, Hepatitis B surface antigen (HbsAg) and the Hepatitis C virus (HCV) antibody and found to be NON-REACTIVE. FDA approved methods have been used to conduct these tests. However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting disease. For *in vitro* diagnostic use only.

STORAGE AND STABILITY

Unopened Lipid Control is stable until the expiry date printed on the product label, when stored between $+2^{\circ}C$ and $+8^{\circ}C$. Once reconstituted, the components of the serum are stable for 7 days at $+2^{\circ}C$ to $+8^{\circ}C$, and 4 weeks at $-20^{\circ}C$ when frozen once. The following exceptions apply: LP(a) is stable for 16 weeks at $-20^{\circ}C$ when frozen once. Values may drop by up to 10% for Direct LDL Cholesterol when stored for 4 weeks at $-20^{\circ}C$.

PREPARATION FOR USE

Open the vial carefully, avoiding any loss of the material and reconstitute with 3 ml of distilled water. Replace the rubber stopper, close the vial and leave to stand for 30 minutes before use. Ensure that all traces of dry material are dissolved by swirling gently.

MATERIALS PROVIDED

Lipid Control - Level I 5×3 ml

MATERIALS REQUIRED BUT NOT PROVIDED

Distilled water

Volumetric pipette

VALUE ASSIGNMENT

Each batch of Lipid Control is submitted to a number of external laboratories. Values are assigned from a consensus of results obtained by these laboratories and internal testing conducted at Randox Laboratories Ltd.

If a method is unavailable, contact Randox Laboratories - Technical Services, Northern Ireland, tel: +44 (0) 28 9445 1070 or email Technical.Services@randox.com.

II Apr'l6 ne

LIPID CONTROL - LEVEL 1 (LPD CONTROL 1) Cat. No. LE2661 Lot No. 2392CH					
Size: 5 x 3 ml Expiry: 2018-10		Range			
Analyte	unit	Target	low	high	methods
Apolipoprotein A-1	g/l	0.99	0.81	1.17	Immunoturbidimetric
	mg/dl	99.3	81.4	117	
	g/l	0.90	0.73	1.06	Nephelometric
	mg/dl	89.5	73.4	106	
Apolipoprotein B	g/I	0.79	0.65	0.93	Immunoturbidimetric
	mg/dl	79.0	64.8	93.2	
	g/l	0.76	0.63	0.90	Nephelometric
	mg/dl	76.3	62.6	90.0	
Cholesterol	mmol/l	3.65	3.18	4.12	Cholesterol Oxidase
	mg/dl	141	123	159	
	mmol/l	3.62	3.15	4.09	Siemens Dimension
	mg/dl	140	122	158	
	mmol/l	3.43	2.98	3.88	Ortho Vitros Microslide Systems
	mg/dl	132	115	149	
HDL - Cholesterol	mmol/l	0.74	0.63	0.85	Direct Clearance Method
	mg/dl	28.6	24.4	32.8	
	mmol/l	0.20	0.14	0.26	Phosphotungstic acid pptn.
	mg/dl	7.64	5.35	9.93	
	mmol/l	0.74	0.63	0.85	Direct HDL Immunoseparation
	mg/dl	28.6	24.4	32.8	
	mmol/l	0.79	0.67	0.91	Direct HDL PEGME
	mg/dl	30.4	25.8	35.0	
	mmol/l	0.84	0.71	0.96	Direct HDL PPD
	mg/dl	32.3	27.4	37.2	
	mmol/l	0.77	0.65	0.88	Direct HDL Roche 3rd generation
	mg/dl	29.6	25.1	34.1	
	mmol/l	0.84	0.72	0.97	Vitros dHDL PTA/MgCl2 direct precipitation
	mg/dl	32.6	27.8	37.4	
	mmol/l	0.89	0.76	1.02	HDL - Ultra
	mg/dl	34.4	29.3	39.5	
LDL - Cholesterol	mmol/l	2.35	2.00	2.70	Direct Clearance Method
	mg/dl	90.7	77.2	104	
	mmol/l	2.34	1.99	2.69	Calculated
	mg/dl	90.3	76.8	104	
	mmol/l	2.25	1.91	2.59	Selective detergent methods
	mg/dl	86.9	73.7	100	
Lipoprotein (a)	mg/dl	13.2	10.6	15.8	Immunoturbidimetric
	nmol/l	27.7	22.2	33.2	
Triglycerides	mmol/l	1.20	1.01	1.39	Lipase/GPO-PAP no correction

mg/dl

106

89.4

123

LIPID CONTROL - LEVEL 1 (LPD CONTROL 1) Cat. No. LE2661 Lot No. 2392CH Size: 5 x 3 ml Expiry: 2018-10 Range **Analyte** unit **Target** low high methods Triglycerides mmol/l 1.25 1.05 1.45 Ortho Vitros Microslide Systems 111 92.9 mg/dl 129 mmol/l 1.20 1.01 1.39 Lipase/GK UV no correction mg/dl 106 89.4 123 mmol/l 1.24 1.04 1.44 Lipase/GPO-PAP 0.11mmol/l correction mg/dl 110 92.0 128

Randox Laboratories Ltd., 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom PAGE 3 OF 3